

MayoComplete Kidney Cancer Panel, Next-Generation Sequencing, Tumor

Overview

Useful For

Identifying specific mutations to assist in tumor diagnosis/classification

Assisting in the clinical management of patients with renal cell carcinoma

Assessment of microsatellite instability for immunotherapy decisions

Genetics Test Information

This test uses targeted next-generation sequencing to determine microsatellite instability (MSI) status and evaluate for somatic mutations within the *ATRX, BAP1, BRAF, CDKN2A, FH, FLCN, KRAS, MET, MITF, MLH1, MSH2, MSH6, MTOR, NF2, PBRM1, PMS2, PTEN, RB1, SDHA, SDHB, SDHC, SDHD, SETD2, SMARCB1, ELOC (TCEB1), TERT, TP53, TSC1, TSC2, and VHL genes.* See <u>Targeted Genes and Methodology Details for MayoComplete Kidney Cancer Panel</u> for details regarding the targeted gene regions evaluated by this test.

This test is performed to evaluate for somatic mutations within solid tumor samples. This test **does not assess** for germline alterations within the genes listed.

Additional Tests

| Test Id | Reporting Name | Available Separately | Always Performed |
|---------|--------------------|----------------------|------------------|
| SLIRV | Slide Review in MG | No, (Bill Only) | Yes |

Testing Algorithm

When this test is ordered, slide review will always be performed at an additional charge.

Special Instructions

- Tissue Requirements for Solid Tumor Next-Generation Sequencing
- Targeted Genes and Methodology Details for MayoComplete Kidney Cancer Panel

Highlights

This panel includes a curated list of 30 genes that are important for the clinical management of patients with renal cell carcinoma. Identifying disease-causing alterations of interest is helpful in establishing the correct diagnostic subtype of renal cell carcinoma and determining prognosis in certain instances.

This test evaluates formalin-fixed, paraffin-embedded tumor or cytology slides from patients with kidney carcinoma for gene mutations to identify candidates for targeted therapy.

This test evaluates mismatch repair genes and microsatellite instability (MSI) status (MSS, MSI-H) as this is often clinically actionable for determining the efficacy of immunotherapy in solid tumors.



MayoComplete Kidney Cancer Panel, Next-Generation Sequencing, Tumor

Method Name

Sequence Capture Next-Generation Sequencing (NGS)

NY State Available

Yes

Specimen

Specimen Type

Varies

Ordering Guidance

Multiple oncology (cancer) gene panels are available. For more information see <u>Hematology, Oncology, and Hereditary</u> <u>Test Selection Guide</u>.

Necessary Information

A pathology report (final or preliminary), at minimum containing the following information, must accompany specimen for testing to be performed:

1. Patient name

2. Block number-must be on all blocks, slides, and paperwork (can be handwritten on the paperwork)

- 3. Tissue collection date
- 4. Source of the tissue

Specimen Required

This assay requires at least 20% tumor nuclei.

-Preferred amount of tumor area with sufficient percent tumor nuclei: tissue 216 mm(2)

-Minimum amount of tumor area: tissue 36 mm(2)

-These amounts are cumulative over up to 10 unstained slides and must have adequate percent tumor nuclei.

-Tissue fixation: 10% neutral buffered formalin, not decalcified

-For specimen preparation guidance, see <u>Tissue Requirement for Solid Tumor Next-Generation Sequencing</u>. In this document, the sizes are given as 4 mm x 4 mm x 10 slides as preferred: approximate/equivalent to 144 mm(2) and the minimum as 3 mm x 1 mm x 10 slides: approximate/equivalent to 36 mm(2).

Preferred:

Specimen Type: Tissue block

Collection Instructions: Submit a formalin-fixed, paraffin-embedded tissue block with acceptable amount of tumor tissue.

Acceptable:

Specimen Type: Tissue slide Slides: 1 Stained and 10 unstained



MayoComplete Kidney Cancer Panel, Next-Generation Sequencing, Tumor

Collection Instructions: Submit 1 slide stained with hematoxylin and eosin and 10 unstained, nonbaked slides with 5-micron thick sections of the tumor tissue.

Note: The total amount of required tumor nuclei can be obtained by scraping up to 10 slides from the same block. **Additional Information**: Unused unstained slides will not be returned.

Specimen Type: Cytology slide (direct smears or ThinPrep)

Slides: 1 to 3 Slides

Collection Instructions: Submit 1 to 3 slides stained and coverslipped with a preferred total of 5000 nucleated cells, or a minimum of at least 3000 nucleated cells.

Note: Glass coverslips are preferred; plastic coverslips are acceptable but will result in longer turnaround times. **Additional Information**: Cytology slides will not be returned.

Specimen Minimum Volume

See Specimen Required

Reject Due To

| Specimens that | Reject |
|------------------|--------|
| have been | |
| decalcified (all | |
| methods) | |
| Specimens that | |
| have not been | |
| formalin-fixed, | |
| paraffin-embe | |
| dded, except | |
| for cytology | |
| slides | |
| Extracted | |
| nucleic acid | |
| (DNA/RNA) | |

Specimen Stability Information

| Specimen Type | Temperature | Time | Special Container |
|---------------|---------------------|------|-------------------|
| Varies | Ambient (preferred) | | |
| | Refrigerated | | |

Clinical & Interpretive

Clinical Information

Molecular genetic profiling is often needed to identify targets amenable to targeted therapies and to minimize treatment costs and therapy-associated risks. Microsatellite instability (MSI) status is an increasingly important



MayoComplete Kidney Cancer Panel, Next-Generation Sequencing, Tumor

biomarker for determining effective immunotherapeutic treatment options for patients with solid tumors.

Renal cell carcinoma is being increasingly subtyped based on underlying molecular alterations. Some kidney tumor types are defined by molecular alterations based on the recent World Health Organization classification of tumors.(1) In addition, the identification of pathognomonic alterations may help classify poorly differentiated tumors, and those associated with hereditary predisposition syndromes. It is important to note that this assay does not distinguish between germline and somatic alterations.

This test uses formalin-fixed paraffin-embedded tissue or cytology slides to assess for somatic mutations involving the following genes known to be associated with kidney cancer: *ATRX, BAP1, BRAF, CDKN2A, FH, FLCN, KRAS, MET, MITF, MLH1, MSH2, MSH6, MTOR, NF2, PBRM1, PMS2, PTEN, RB1, SDHA, SDHB, SDHC, SDHD, SETD2, SMARCB1, ELOC (TCEB1), TERT, TP53, TSC1, TSC2, and VHL, as well as MSI status.* The results of this test can be useful for assessing prognosis and guiding treatment of individuals with kidney tumors. These data can also be used to help determine clinical trial eligibility for patients with alterations in genes not amenable to current US Food and Drug Administration-approved targeted therapies.

Reference Values

An interpretive report will be provided.

Interpretation

The interpretation of molecular biomarker analysis includes an overview of the results and the associated diagnostic, prognostic, and therapeutic implications.

Cautions

This test cannot differentiate between somatic and germline alterations. Additional testing may be necessary to clarify the significance of results if there is a potential hereditary risk.

DNA variants of uncertain significance may be identified.

A negative result does not rule out the presence of a variant that may be present below the limits of detection of this assay. In a sample with 20% or more tumor content, the analytical sensitivity of this assay for sequence reportable alterations is 5% mutant allele frequency with a minimum coverage of 500X.

Point mutations and small deletion-insertion mutations will be detected in the ATRX, BAP1, BRAF, CDKN2A, FH, FLCN, KRAS, MET, MITF (exon 9 only), MLH1, MSH2, MSH6, MTOR, NF2, PBRM1, PMS2, PTEN, RB1, SDHA, SDHB, SDHC, SDHD, SETD2, SMARCB1, ELOC (TCEB1), TERT(5'UTR), TP53, TSC1, TSC2, and VHL genes only. This test may detect single exon deletions but does not detect multi-exon deletions, duplications, or genomic copy number variants (such as chromosome 3p deletions).

Variant allele frequency (VAF) is the percentage of sequencing reads supporting a specific variant divided by the total sequencing reads at that position. In somatic testing, VAF should be interpreted in the context of several factors including, but not limited to, tumor purity/heterogeneity/copy number status (ploidy, gains/losses, loss of heterozygosity) and sequencing artifact/misalignment.(2,3)

Rare alterations (ie, polymorphisms) may be present that could lead to false-negative or false-positive results.



MayoComplete Kidney Cancer Panel, Next-Generation Sequencing, Tumor

The presence or absence of a variant may not be predictive of response to therapy in all patients.

Test results should be interpreted in the context of clinical, tumor sampling, histopathological, and other laboratory data. If results obtained do not match other clinical or laboratory findings, contact the laboratory for discussion. Misinterpretation of results may occur if the information provided is inaccurate or incomplete.

Reliable results are dependent on adequate specimen collection and processing. This test has been validated on cytology slides and formalin-fixed, paraffin-embedded tissues; other types of fixatives are discouraged. Improper treatment of tissues, such as decalcification, may cause polymerase chain reaction failure.

Supportive Data

Performance Characteristics

The limit of detection for calling a somatic variant (single nucleotide variants [SNV] and deletions/insertions [delins, formerly indel]) is 5% variant allele frequency (VAF) and having at least 500x deduplicated coverage.

Verification studies demonstrated concordance between this test and the reference method for detection of SNV and delins is 98.5% (673/683) and 98.4% (122/124) of variants, respectively. Concordance for the detection of delins was 99.0% (100/101) in variants 1 to 10 base pairs (bp) in size, 93.3% (14/15) in variants 11 to 50 bp in size, and 100% (8/8) in variants over 50 bp in size.

Microsatellite instability (MSI) evaluation is accurate at a tumor purity of at least 10% for colorectal tumors and 20% for other tumor types. During verification studies, 98% (200/204) concordance for MSI status was observed between this test and the reference method.

To ensure accuracy, this test will be performed on cases that are estimated by a pathologist to have at least 20% tumor cells.

Clinical Reference

1. WHO Classification of Tumours Editorial Board, eds. Urinary and male genital tumors. 5th ed. World Health Organization; 2022. WHO Classification of Tumours. Vol 8

2. Strom SP. Current practices and guidelines for clinical next-generation sequencing oncology testing. Cancer Biol Med. 2016;13(1):3-11. doi:10.28092/j.issn.2095-3941.2016.0004

3. Spurr L, Li M, Alomran N, et al. Systematic pan-cancer analysis of somatic allele frequency. Sci Rep. 2018;8(1):7735. Published 2018 May 16. doi:10.1038/s41598-018-25462-0

4. Trpkov K, Hes O, Williamson SR, et al: New developments in existing WHO entities and evolving molecular concepts: The Genitourinary Pathology Society (GUPS) update on renal neoplasia. Mod Pathol. 2021;34(7):1392-1424

5. Trpkov K, Williamson SR, Gill AJ, et al. Novel, emerging and provisional renal entities: The Genitourinary Pathology Society (GUPS) update on renal neoplasia. Mod Pathol. 2021;34(6):1167-1184

6. Marcus L, Lemery SJ, Keegan P, Pazdur R: FDA Approval Summary: Pembrolizumab for the treatment of microsatellite instability-high solid tumors. Clin Cancer Res. 2019;25(13):3753-3758

Performance



MayoComplete Kidney Cancer Panel, Next-Generation Sequencing, Tumor

Method Description

Next-generation sequencing is performed to determine microsatellite instability (MSI) status and evaluate the presence of a mutation in most coding regions of the ATRX, BAP1, BRAF, CDKN2A, FH, FLCN, KRAS, MET, MITF, MLH1, MSH2, MSH6, MTOR, NF2, PBRM1, PMS2, PTEN, RB1, SDHA, SDHB, SDHC, SDHD, SETD2, SMARCB1, ELOC (TCEB1), TERT, TP53, TSC1, TSC2, and VHL genes. See <u>Targeted Genes and Methodology Details for MayoComplete Kidney Cancer Panel</u> for details regarding the targeted gene regions identified by this test.(Unpublished Mayo method)

A pathology review and macro dissection to enrich for tumor cells is performed prior to slide scraping.

PDF Report

No

Day(s) Performed

Monday through Friday

Report Available

12 to 20 days

Specimen Retention Time

FFPE tissue block: Unused portions of blocks will be returned 10 to 14 days after testing is complete; FFPE tissue/cytology slides: Unused slides are stored indefinitely; Digital images are obtained and stored for all slides used in testing

Performing Laboratory Location

Rochester

Fees & Codes

Fees

- Authorized users can sign in to <u>Test Prices</u> for detailed fee information.
- Clients without access to Test Prices can contact <u>Customer Service</u> 24 hours a day, seven days a week.
- Prospective clients should contact their account representative. For assistance, contact <u>Customer Service</u>.

Test Classification

This test was developed and its performance characteristics determined by Mayo Clinic in a manner consistent with CLIA requirements. It has not been cleared or approved by the US Food and Drug Administration.

CPT Code Information

88381-Microdissection, manual 81457

LOINC[®] Information

Document generated July 22, 2024 at 02:32 AM CT

Test Definition: MCKCP

MayoComplete Kidney Cancer Panel, Next-Generation Sequencing, Tumor

| Test ID | Test Order Name | Order LOINC [®] Value |
|-----------|----------------------------------|---------------------------------|
| МСКСР | MayoComplete Kidney Cancer Panel | In Process |
| | | |
| Result ID | Test Result Name | Result LOINC [®] Value |
| 619596 | Result | 82939-0 |
| 619597 | Interpretation | 69047-9 |
| 619598 | Additional Information | 48767-8 |
| 619599 | Specimen | 31208-2 |
| 619600 | Tissue ID | 80398-1 |
| 619601 | Method | 85069-3 |
| 619602 | Disclaimer | 62364-5 |
| 619603 | Released By | 18771-6 |

